

M200-DI FAQs

Q: Can I use the M200-DI with wet gasses?

A: Potentially corrosive or wet gasses can be handled by the M200-DI.

Q: Can I order an Intrinsic Safety Certification?

A: Yes.

Q: Can the M201-DI be recalibrated in the field?

A: Yes. The M200-DI can be recalibrated in the field for zero, span and linearity. The user has the option to select a 1 point or 5 point calibration procedure. Enables users with accurate primary standards to calibrate unit locally if desired.