

Addendum to M200 Series User Manual No. 9R000045

Software improvements:

M200	M200LS
1. When the measurement is outside of the calibrated range of pressure and/or temperature, the red backlight automatically displays.	1. When the measurement is outside of the calibrated range of pressure and/or temperature, the red backlight automatically displays.
2. N/A	2. The red backlight automatically displays when out of the temperature range for 0.01%, indicating the default accuracy of 0.025%.
3. When the measurement is well outside of the calibrated range, an error message automatically displays in place of reading.	3. When the measurement is well outside of the calibrated range, an error message automatically displays in place of reading.
NOTE: The information here replaces the accuracy statement in the first paragraph in Appendix 3.	